

Invader[®] MTHFR 677 510(k) SUMMARY

MAY 13 2011

A. 510(k) Number:

K100987

B. Purpose for Submission:

New Device

C. Measurand:

MTHFR 677

D. Type of Test:

Qualitative genotyping test for single nucleotide polymorphism detection.

E. Applicant:

Hologic Inc.

Third Wave Technologies

250 Campus Drive

Marlborough, MA 01752

508-263-8912

Contact Person: Randall J. Covill, Manager, Regulatory Affairs

Date of Submission: April 2010

F. Proprietary and Established Names:

Invader[®] MTHFR 677

G. Regulatory Information:

1. Regulation Sections: 21 CFR 864.7280

2. Classification:

Class II

3. Product Code:

OMM: Test 5,10-Methylenetetrahydrofolate Reductase Mutations, Genomic DNA PCR

4. Panel:

Hematology (81)

H. Intended Use:

1. Intended Use(s):

The Invader[®] MTHFR 677 test is an *in vitro* diagnostic test intended for the detection and genotyping of a single point mutation (C to T at position 677) of the human 5,10-methylenetetrahydrofolate reductase (MTHFR) gene in isolated genomic DNA obtained from whole blood Potassium EDTA samples from patients with suspected thrombophilia.

2. Indication(s) for use:

Same as Intended Use

3. Special Conditions for use statements(s):

For prescription use only

4. Special instrument requirements:

None

I. Device Description:

The Invader MTHFR 677 test consists of the following components:

MTHFR 677 Oligo Mix

Universal Buffer

Universal Enzyme Mix

No DNA Control

MTHFR 677 Wild Type Control

MTHFR 677 Heterozygous Control

MTHFR 677 Mutant Control

Invader Call Reporter[™] Software

J. Substantial Equivalence Information:

1. Predicate device name(s):
Verigene® MTHFR Nucleic Acid Test (Nanosphere, K070597)
2. Predicate 510(k) number(s):
Nanosphere, K070597
3. Comparison with predicate:

Table 1: Comparison with Predicate Device

	<i>Predicate Device</i>	<i>Proposed Device</i>
Product Name (Manufacturer, Submission)	Verigene® MTHFR Nucleic Acid Test (Nanosphere, K070597)	Invader® MTHFR 677 (Hologic, Inc. K100987)
Intended Use	<p>"The Verigene MTHFR Nucleic Acid Test is an <i>in vitro</i> diagnostic for the detection and genotyping of a single point mutation (C to T at position 677) of the human 5, 10 methylene-tetrahydrofolate reductase gene (MTHFR) in patients with suspected thrombophilia, from isolated genomic DNA obtained from whole blood samples. The test is intended to be used on the Verigene System."</p> <p>The Verigene System is a bench-top molecular diagnostic workstation that automates the <i>in vitro</i> diagnostic analysis and detection of nucleic acids using gold nanoparticle probe technology. The Verigene System is intended to be used by experienced laboratory professionals with training on basic laboratory techniques and on the use of the system components.</p>	The Invader® MTHFR 677 test is an <i>in vitro</i> diagnostic test intended for the detection and genotyping of a single point mutation (C to T at position 677) of the human 5,10-methylenetetrahydrofolate reductase (MTHFR) gene in isolated genomic DNA obtained from whole blood Potassium EDTA samples from patients with suspected thrombophilia.
Specimen Type	Purified DNA isolated from human whole peripheral blood	Same as predicate
Indications for	Same as Intended Use	Same as Intended Use

Use		
Target Population	Patients with suspected thrombophilia	Same as predicate
Chemistry	SNP discrimination via oligonucleotide probes; detection via evanescent wave light scatter with nanoparticles.	PCR followed by Invader® achieving SNP discrimination via oligonucleotide probes.
Hardware	The Verigene System consists of two instruments, the Verigene Processor and the Verigene Reader, and utilizes single-use, disposable Test Cartridges to process and genotype multiple genes in a DNA samples in approximately 1.5 hours.	Non-specified, third-party fluorometer and thermal cyclor.
Software Interface	Embedded software in closed system, integrated graphical user interface.	Java-based software installed on a standalone PC capable of converting raw fluorescence data into genotype calls.
Detection Method	Single-image sensor where nanoparticles are illuminated using a fixed-wavelength light source.	PCR and Fluorescence Resonance Energy Transfer (FRET) chemistry for signal reporting.
Sample Size	25µL	20ul reaction containing 0.25-4ng/ul gDNA extracted from human peripheral whole blood.
Detection Procedure	Single-image sensor where nanoparticles are illuminated using a fixed-wavelength light source.	Multi-well fluorometer to detect raw fluorescence.
Detection Chemistry	Detection via evanescent wave light scatter with nanoparticles.	PCR and Invader® using Fluorescence Resonance Energy Transfer (FRET) chemistry for signal reporting.
Analysis Time	90 min. processing with 2 min analysis time.	~90 min. amplification followed by 1 min signal detection. Software analysis post signal detection.

K. Standard/Guidance Document Referenced (if applicable):

- Guidance for Industry and FDA Staff – Class II Special Controls Guidance Document: Factor V Leiden DNA Mutation Detection Systems issued on March 16, 2004
- Guidance for Industry and FDA Staff - Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices issued May 11, 2005
- Guidance for Industry and FDA Staff – Format for Traditional and Abbreviated 510(k)s issued on August 12, 2005

L. Test Principle:

The Invader[®] MTHFR 677 test utilizes the Invader Plus[®] chemistry with DNA isolated from human whole blood, for the detection of the targeted sequence polymorphism. Specifically, the Invader Plus[®] chemistry utilizes a single-tube, two phase reaction, including target amplification and signal generation (mediated by Invader[®] chemistry). Invader Plus[®] reaction mixes are assembled by combining the MTHFR 677 Oligo Mix, Universal Enzyme Mix, and Universal Buffer. In a 96-well plate, reaction mix is combined with purified genomic DNA samples, as well as four (4) controls included with the test. The No DNA Control is used by the interpretive software to set the "noise" component of the run for "signal-to-noise" calculations. The genotype-specific controls (WT, HET, MUT) ensure reagents were assembled correctly and perform according to the specifications. The 96-well plate is transferred to an appropriately programmed thermal cycler for target amplification and signal generation. In the target amplification phase of the reaction, amplification is carried out using "two-step" cycling conditions (i.e. denaturation & annealing/extension). Following amplification, Taq polymerase is inactivated by a 10 minute incubation at 99°C, after which the thermal cycler proceeds to 63°C to initiate the signal generation (Invader[®]) phase of the reaction (see Figure 1).




Figure 1. Invader® Signal Generation Phase

During the signal generation phase, a discriminatory Primary Probe transiently hybridizes to the amplified target sequence along with an Invader® oligonucleotide, to form an overlapping structure. The 5'-end of the Primary Probe includes a 5'-flap that does not hybridize to the target DNA. The 3'-nucleotide of the bound Invader® oligonucleotide overlaps the Primary Probe, and does not hybridize to the target DNA. The Cleavase® enzyme recognizes this overlapping structure and cleaves off the unpaired 5'-flap of the Primary Probe, releasing it as a target-specific product. The Primary Probe is designed to have a melting temperature aligned with the Invader® reaction temperature so that under the isothermal reaction conditions (~63°C) the Primary Probes cycle on and off the target DNA. This allows for multiple rounds of Primary Probe cleavage for each DNA target resulting in an accumulation of the number of released 5'-flaps. The released 5'-flap transiently hybridizes with a corresponding FRET cassette forming an overlapping structure that is recognized and the fluorophore is cleaved from the FRET cassette by the Cleavase® enzyme. The 5'-flap is designed to have a melting temperature aligned with the Invader® reaction temperature, so that the 5'-flaps cycle on and off of the corresponding FRET cassettes. This allows for multiple rounds of FRET cassette cleavage for each 5'-flap, and an accumulation of released fluorophore. When the FRET cassette is cleaved, a fluorophore and quencher are separated, generating detectable fluorescence signal. The format uses two different discriminatory Primary Probes, one for the mutant allele and one for the wild type allele (Figure 1). Each Primary Probe is assigned a unique 5'-flap, and distinct FRET cassette, with a spectrally distinct fluorophore. By design, the released 5'-flaps will bind only to their respective FRET cassettes to generate a target-specific signal, linking the wild type allele with one fluorophore (Fluorescence 1: FAM) and the mutant allele with the second fluorophore (Fluorescence 2: RED).

The Invader® MTHFR 677 software, in combination with Invader Call Reporter™ software, is a data analysis software package developed by Hologic for use with the Invader® MTHFR 677 test. The software package provides a working template for the setup of reaction mixes and sample placement, and following the import of fluorescence data, it determines results and validity for controls and samples. A summary of the Invader Call Reporter™- Invader® MTHFR 677 package workflow is shown in Figure 2.




Figure 2. Invader Call Reporter™- Invader® MTHFR 677 Package Workflow

M. Performance Characteristics (if/when applicable):

1. Analytical performance:

a. Precision/Reproducibility:

External Reproducibility (Study #1): Two operators each from three (3) different sites (2 external sites and 1 internal site) performed the testing, in duplicate, over five (5) non-consecutive days for a ten (10) day period using the same testing materials including a panel of nine (9) whole blood samples specific for each of the three (3) possible genotypes (i.e. wild type, heterozygous, homozygous mutant).

Table 2: Inter-laboratory Reproducibility of Invader® 677 Test									
Site	Operator	Samples tested	First Pass			Final			Final % Agreement
			Correct Calls	No Calls (Invalid, EQ)	Miscalls	Correct Calls	No Calls (Invalid, EQ)	Miscalls	
Site 001	1	90	90	0	0	90	0	0	100%
	2	90	90	0	0	90	0	0	100%
Site 002	1	90	90	0	0	90	0	0	100%
	2	90	90	0	0	90	0	0	100%
Site 003	1	90	90	0	0	90	0	0	100%
	2	90	90	0	0	90	0	0	100%
All	All	540	540	0	0	540	0	0	100%

Lot-to-Lot Reproducibility (Study #9): A total of nine (9) genomic DNA samples (three (3) wild type, three (3) heterozygous and three (3) mutants) were tested in quadruplicate using three (3) different kit lots of the Invader® MTHFR 677 test. The percent agreement between Invader® MTHFR 677 test and sequencing was 100% (n=108).

Table 3: Lot to Lot Reproducibility						
Lot	# Samples Tested	First Pass Correct Calls	First Pass No Calls	Miscalls	Final Correct Calls	Final Agreement %
1	36	36	0	0	36	100
2	36	36	0	0	36	100
3	36	36	0	0	36	100
Total	108	108	0	0	108	100

b. Linearity/assay reportable range:

Refer to paragraph D below.

c. Traceability, Stability, Expected values (controls, calibrators, or methods):

Real-Time Stability Study (Study #5): Three (3) lots of product in the final configuration are being stored under recommended conditions: (1) -30° to -15°C (Standard Storage of intermediate components) as well as (2) +2° to +8°C (Standard Storage of Genotype-Specific Controls). Functional testing is performed with samples representing all 3 genotypes in quadruplicate at each time point. The interim test results have demonstrated 7 months stability for the device.

Table 4: MTHFR 677 Genotype Results; Real-time Stability

Sample/ Control	Sequencing/ Expected MTHFR 677 Genotype	T ₀ Result			T ₁ Result			T ₂ Result		
		Lot 1	Lot 2	Lot 3	Lot 1	Lot 2	Lot 3	Lot 1	Lot 2	Lot 3
Control 1	WT	WT	WT	WT	WT	WT	WT	WT	WT	WT
Control 2	HET	HET	HET	HET	HET	HET	HET	HET	HET	HET
Control 3	MUT	MUT	MUT	MUT	MUT	MUT	MUT	MUT	MUT	MUT
gDNA 1	WT	WT	WT	WT	WT	WT	WT	WT	WT	WT
gDNA 2	WT	WT	WT	WT	WT	WT	WT	WT	WT	WT
gDNA 3	HET	HET	HET	HET	HET	HET	HET	HET	HET	HET
gDNA 4	MUT	MUT	MUT	MUT	MUT	MUT	MUT	MUT	MUT	MUT
Percent Agreement		100	100	100	100	100	100	100	100	100

Reagent Freeze-Thaw Stability Study (Study #6): Product in the final configuration was subject to 15 freeze-thaw cycles prior to the final thaw at the time of testing. Functional testing was performed using genomic DNA isolated from cell lines, representing all possible genotypes. The percent agreement between the sequencing result and the Invader® MTHFR 677 test were 100%, therefore demonstrating stability for up to fifteen (15) freeze/thaw cycles.

Table 5: Freeze/Thaw Stability of Invader® MTHFR 677

Sample	Number of Freeze/Thaw Cycles															Total	% Agreement
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15		
Control 1 (WT)	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	45	100
Control 2 (HET)	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	45	100
Control 3 (MUT)	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	45	100
gDNA (WT)	6	*	6	*	6	*	*	*	*	6	*	6	*	*	6	36	100
gDNA (HET)	8	*	8	*	8	*	*	*	*	8	*	8	*	*	8	48	100
gDNA (MUT)	6	*	6	*	6	*	*	*	*	6	*	6	*	*	6	36	100
Total	29	9	29	9	29	9	9	9	9	29	9	29	9	9	29	255	100

*Testing with gDNA samples did not occur at this testing point.

d. Detection limit/Analytical Sensitivity and Normal Range (Study #3):

Three (3) genomic DNA samples with different genotypes (i.e. WT, HET, MUT) were extracted from whole blood collected in potassium EDTA. Each sample was diluted to eight different concentrations 0.5, 5, 20, 40, 80, 200, 400, 800 ng/μL and tested in replicates of forty (40). The recommend range of the assay was

determined to be between 5-80 ng/μL of input gDNA, based on 100% concordance of all tested replicates with bi-directional sequencing.

Table 6: Percent Agreement Between Replicates			
Sample ID (Genotype based on Sequencing)			
Input Sample Concentration	03-4520 (MUT)	03-4525 (HET)	03-4524 (WT)
0.5 ng/μl	100% (40/40)	100% (40/40)	100% (40/40)
5 ng/μl	100% (40/40)	100% (40/40)	100% (40/40)
20 ng/μl	100% (40/40)	100% (40/40)	100% (40/40)
40 ng/μl	100% (40/40)	100% (40/40)	100% (40/40)
80 ng/μl	100% (40/40)	100% (40/40)	100% (40/40)
200 ng/μl	100% (40/40)	100% (40/40)	100% (40/40)
400 ng/μl	100% (40/40)	100% (40/40)	100% (40/40)
800 ng/μl	100% (40/40)	100% (40/40)	100% (40/40)

e. Analytical specificity (Interfering Substances) (Study #4):

Test performance was not affected by addition of the following substances to nine (9) whole blood samples of different genotype (3 WT, 3 HET, 3 MUT) prior to extraction:

- Heparin (1500 U/dL human whole blood)
- Cholesterol (300 mg/dL human whole blood)
- Bilirubin (10 mg/dL human whole blood)
- Hemoglobin (up to 0.2% in whole blood)
- Potassium EDTA (1.8 mg/mL human whole blood)
- Ethanol-based Wash Buffer (5% in DNA sample)

Table 7: Summary, Comparison of Invader® MTHFR 677 Interfering Substance Results to Sequencing				
Interfering Substance Code	Substance Concentration / (in blood or DNA sample)	% Agreement with Sequencing Genotype	% Agreement with Untreated Sample Invader® MTHFR 677 Genotype	PASS / FAIL
A	No Addition Control (Untreated)	100% (18 of 18)	N/A	PASS
B	Bilirubin 10mg/dl (blood)	100% (18 of 18)	100% (18 of 18)	PASS
C	Cholesterol 300mg/dl (blood)	100% (18 of 18)	100% (18 of 18)	PASS
D	K ₂ EDTA 1.8mg/ml (blood)	100% (18 of 18)	100% (18 of 18)	PASS
E	Heparin 1500 U/dl (blood)	100% (18 of 18)	100% (18 of 18)	PASS
F1	Hemoglobin 0.2% (blood)	100% (18 of 18)	100% (18 of 18)	PASS
F2	Hemoglobin 0.1% (blood)	100% (18 of 18)	100% (18 of 18)	PASS
F3	Hemoglobin 0.05% (blood)	100% (18 of 18)	100% (18 of 18)	PASS
F4	Hemoglobin 0.025% (blood)	100% (18 of 18)	100% (18 of 18)	PASS
G	Buffer AW2 5% (DNA)	100% (18 of 18)	100% (18 of 18)	PASS

- f. Pre-Analytical Equivalency Study/Genomic DNA Extraction Reproducibility (Study #7): Thirty (30) human whole blood samples and ten (10) leukocyte depleted whole blood spiked with cell lines were divided and extracted using four (4), commercially available DNA extraction methods (A. Qiagen QIAamp® 96 DNA Blood Kit, B. Qiagen QIAamp® DNA Blood Mini Kit, C. Gentra Generation® Capture Column Kit (Qiagen), D. Roche MagNA Pure LC DNA Isolation Kit I). The 160 extracted DNAs were analyzed in singlicate with one (1) lot of the device. The percent agreement between the Invader® MTHFR 677 test for each extraction method and bi-directional sequencing was 100% (n=40).

Table 8a: Pre-Analytical Equivalency						
Extraction Method	# Samples Tested	First Pass Correct Calls	First Pass No Calls	Miscalls	Final Correct Calls	Final Agreement %
A	40	40	0	0	40	100
B	40	39	1*	0	39*	100*
C	40	40	0	0	40	100
D	40	40	0	0	40	100
Total	160	159	1	0	159	100
*Sample was removed from study due to loss of traceability of the sample identification.						

Table 8b: Genotype Distribution Across Sample Type				
Sample Type	WT	MUT	HET	Total
Human Whole Blood	10	9*	10	29*
LDWB Spiked with Cell Lines**	5	0	5	10
Total	15	9*	15	39*
*Sample was removed from study due to loss of traceability of the sample identification.				
**LDWB=Leukocyte Depleted Whole Blood				

- g. Instrument Equivalency (Study #8): Twenty-nine (29) human whole blood samples and ten (10) leukocyte depleted whole blood samples spiked with cell lines were extracted using two (2) commonly used extraction methods. The extracts were tested with the Invader® MTHFR 677 test using three (3) commercially available thermal cyclers (1. ABI GeneAmp® PCR System 9700 with 96-well gold block, 2. ABI Veriti™ and 3. MJ Research PTC-100) and the raw fluorescent data acquired on three (3) commercially available fluorometers (A. Tecan Infinite®, B. Tecan Genios® and C. BioTek®, FLx800). Results from the three (3) fluorometers were transferred into the interpretive software and genotype calls compared to bi-directional sequencing.

Table 9: Concordance by Instrument			
Fluorometer	Thermal Cycler		
	1	2	3
A	78 of 78 = 100%	78 of 78 = 100%	78 of 78 = 100%
B	78 of 78 = 100%	78 of 78 = 100%	78 of 78 = 100%
C	78 of 78 = 100%	78 of 78 = 100%	78 of 78 = 100%

2. Comparison studies:

a. Method comparison: Bi-directional Sequencing (Study #2):

Human whole blood samples (n = 361) underwent DNA extraction and subsequent bi-directional DNA sequence analysis. The same DNA samples were then analyzed using the Invader® MTHFR 677 test. The observed agreement between the Invader® MTHFR 677 test and bi-directional DNA sequencing was 100% (359/359). The first run agreement with bi-directional sequencing was 99.45% (359/361).

Table 10a: Agreement between the Invader® MTHFR 677 Test and Bi-directional DNA Sequencing

MTHFR 677 Genotype*	Number tested	Number of valid results on 1 st run	Number of Correct genotype calls on First Run	First Run Agreement
Homozygous Wild Type (GG)	180	178**	178**	98.89%
Heterozygous (GA)	104	104	104	100%
Homozygous Mutant (AA)	77	77	77	100%
Total	361	359**	359**	99.45%

* Genotype determined through bi-directional DNA sequencing

** Two samples failed to generate valid results. These samples were reported as invalid (EQ) and no genotype calls were assigned by the interpretive software. The EQ result was used to determine the First Run Agreement.

Table 10b: Agreement between the Invader® MTHFR 677 Test and Bi-directional DNA Sequencing

Genotype*						First Pass Results		Genotype					Final Results	
	WT	HET	MUT	No-Call	Mis-Call	% Agreement†	95% LCB**	WT	HET	MUT	No-Call	Mis-Call	% Agreement	95% LCB
WT	178	0	0	2	0	99.45	96.04	178	0	0	2	0	99.45	96.04
HET	0	104	0	0	0	100.00	97.16	0	104	0	0	0	100.00	97.16
MUT	0	0	77	0	0	100.00	96.18	0	0	77	0	0	100.00	96.18

*Genotype determined through bi-directional DNA sequencing. **Lower boundary of the 95% confidence interval for each genotype. †Calculated across all samples.

3. External Reproducibility studies:
 - a. *Clinical Sensitivity:* please refer to section 1d above.
 - b. *Clinical specificity:* please refer to section 1e above.
4. Expected values/Reference range: (Prevalence)
MTHFR 677: 30-40% Caucasians

N. System Descriptions:

1. Modes of Operation:
Closed System

2. Software:
FDA has reviewed applicant's Hazard Analysis and software development processes for this line of product type. Yes ___X___ or No _____
3. Specimen Identification:
Manual Labeling
4. Specimen Sampling and Handling:
DNA should be extracted using a validated DNA extraction method that generates DNA concentration range of greater than 5ng/μl.
5. Quality Control:
Each test contains positive and negative controls to assure proper functioning of the system: Failure of any controls will be indicated as "Invalid" in the test results section of the report. The genotyping test result will not be reported for any sample for which a positive or negative control failure occurs.
Positive Control: The genotype controls (WT, HET, MUT) ensure reagents were assembled correctly and perform according to the specifications.
Negative Control: The No DNA Control is used by the interpretive software to set the "noise" component of the run for "signal-to-noise" calculations.
Hardware and Software Controls:
The genotyping test result will not be reported for any sample for which a positive or negative control failure occurs.

O. Proposed Labeling:

The labeling is sufficient and satisfies the requirements of 21 CFR Part 809.10.

P. Conclusion:

The submitted information in this 510 (k) notification is complete and supports a substantial equivalence decision.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Food and Drug Administration
10903 New Hampshire Avenue
Document Mail Center – WO66-0609
Silver Spring, MD 20993-0002

Hologic Inc.
Third Wave Technologies
c/o Mr. Randall J. Covill
Senior Specialist Regulatory Affairs
502 South Rosa Road
Madison, WI 53719

MAY 13 2011

Re: k100987

Trade/Device Name: Invader® MTHFR 677
Regulation Number: 21 CFR §864.7280
Regulation Name: Factor V Leiden DNA Mutation Detection Systems
Regulatory Class: Class II
Product Code: OMM
Dated: March 31, 2011
Received: April 7, 2011

Dear Mr. Covill:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into class II (Special Controls), it may be subject to such additional controls. Existing major regulations affecting your device can be found in Title 21, Code of Federal Regulations (CFR), Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); and good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820). This letter


Page 2 – Mr. Randall J. Covill

will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Parts 801 and 809), please contact the Office of *In Vitro* Diagnostic Device Evaluation and Safety at (301) 796-5450. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Maria M. Chan, Ph.D
Director
Division of Immunology and Hematology Devices
Office of *In Vitro* Diagnostic Device Evaluation and Safety
Center for Devices and Radiological Health

Enclosure

Indication for Use Statement

510(k) Number (if known): k100987

Device Name: Invader MTHFR 677 test

Indication for Use:

The Invader® MTHFR 677 test is an in vitro diagnostic test intended for the detection and genotyping of a single point mutation (C to T at position 677) of the human 5,10-methylenetetrahydrofolate reductase (MTHFR) gene in isolated genomic DNA obtained from whole blood potassium EDTA samples from patients with suspected thrombophilia.

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE OF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)



Division Sign-Off

Office of In Vitro Diagnostic
Device Evaluation and Safety

510(k) k100987